



1/4

Sheet No.

GT200-OF032E Oil

Determination of iodine number of palm oil (Wijs-Cyclohexane method)

Method : Oxidation-reduction titration

Apparatus : Automatic Titrator model GT-200 (GT0EF)

Electrodes: Combined platinum electrode (GTPR1B)

*Inner solution: 3.3mol/I Potassium chloride

Titration mode : INF, Detection: pH / mV

Related standard : Japan Oil Chemists' Society "JOCS Standard Methods for the Analysis of Fats, Oils and

Related Materials," "Iodine Numbers (Wijs-Cyclohexane Method)"

*This application sheet is provided as reference, and does not assure the measurement results. Please consider analysis environment, external factors and sample nature foroptimal conditions before the measurement

Outline

Palm oil is a vegetable oil obtained from the fruit of an oil palm. In addition to the use as cooking oil, it is used as raw materials for soap, detergent and biodiesel. Iodine numbers are measured to confirm the number of carbon-carbon double bonds (C=C) contained in a sample, and expressed by converting the halogen content which reacts with 100g sample into the number of grams of iodine.

Reagents

[Titrant]

■0.1mol/L-sodium thiosulfate in water (Volumetric analysis grade)

[Reagents]

- ■Cyclohexane (Special grade)
- ■Wijs solution (0.1mol/L-iodine chloride solution in acetic acid (For measurement of iodine numbers)) [Adjustment reagents]
- ■10g/100ml potassium iodide solution ... Dissolve 10g potassium iodide in pure water to 100ml total.

Analytical Procedure

[Pretreatment]

As palm oil is solid at room temperature, melt it down in a water bath before the determination.

[Measurement]

- (1) Add 2g melted sample into a 200ml beaker.
- (2) Collect 10ml cyclohexane using a Komagome pipette and add it into the above-mentioned beaker to dissolve the sample.
- (3) Collect 25ml Wijs solution using a whole pipette and add it into the beaker.
- (4) Leave the beaker still in a dark place for 30 minutes shaking it occasionally.
- (5) After leaving it still, collect 20ml of 10g/100ml potassium iodide solution using a whole pipette and add it into the beaker.
- (6) Collect 100ml pure water using a measuring cylinder and add it into the beaker. Titrate with 0.1mol/L-sodium thiosulfate solution while agitating strongly.
- (7) Perform blank measurements in the same way.



Sheet No.

GT200-OF032E Determination of iodine number of palm oil (Wijs-Cyclohexane method) 2/4

[Calculation]

lodine number $(g/100g) = (BL - A1) \times f \times 1.269/S$

BL: Titration volume of 0.1mol/L-sodium thiosulfate solution at blank measurement (ml): Titration volume of 0.1mol/L-sodium thiosulfate solution at sample titration (ml)

f : Factor of 0.1mol/L-sodium thiosulfate solution

1.269 : 0.01269 (The number of grams of iodine corresponding to 1ml of 0.1mol/L-sodium thiosulfate

solution) . 100. (Converted into the value per 100g)

S : Sample volume (g)

Other Requirements

■Make sure to confirm labels and safety data sheets of reagents and gases used for the measurement and handle them with enough care.

■Wear protective equipment (eye protector, gloves and others) when handling reagents.

Measurement Results

| | Sample size (g) | Titration volume (ml) | Blank (ml) | Results (g/100g) |
|---|-----------------|-----------------------|------------|------------------|
| 1 | 2.0174 | 1.3199 | 46.0716 | 28.89 |
| 2 | 2.0221 | 1.3691 | 46.0224 | 28.80 |
| 3 | 2.0038 | 1.3367 | 46.0548 | 29.08 |

Nos. of data (n) 3
Average 28.92
Standard deviation (SD) 0.14
Relative standard deviation (RSD%) 0.50

Blank 47.3915 ml

lodine numbers of palm oil were measured using GT-200. The average of three measurements was 28.9g/100g and the relative standard deviation (RSD %) was 0.50%. GT-200 can measure iodine numbers with good repeatability.


★ MITSUBISHI CHEMICAL ANALYTECH



Sheet No.

GT200-OF032E Determination of iodine number of palm oil (Wijs-Cyclohexane method) 3/4

ID No.: 8 GT No.1 User: GT-200



Run file No.: 17

Titration file No.: 58 palm oil lodine blank

*Run file and Titration file parameters are set for each analysis item

Mode : INF End1, End1 Width : $500 [mV] \pm 1500 [mV]$

Detect : mV1

BRT No. : 1 Preset 1 Mode: V BRT: 1

Reagent : 32 Volume : 35 [ml] WTint : 0 [sec] Position : Titration

Vup : 300 [µl]
Vlow : 20 [µl]
dE : 5 [mV]
dT : 5 [sec]

DL : 600 [mV/ml]

DetCnt : 10 Vmax : 50 [ml]

Vover : 1 [ml]

Reagent name (Reag): Na2S2O3 Equivalent (E): 1 Molarity(M): 0.1 [Mol/I]

C1:A1

Factor (f) : 0.997

Buret Injection Speed: 500 [ul/sec]


★ MITSUBISHI CHEMICAL ANALYTECH



Sheet No.

GT200-OF032E Determination of iodine number of palm oil (Wijs-Cyclohexane method) 4/4

ID No.: 5 GT No.1 User: GT-200



Run file No.: 17

Titration file No.:59 Iodine number of palm oil

*Run file and Titration file parameters are set for each analysis item

Mode : INF End1, End1 Width: 500 [mV] ± 1500 [mV]

Detect: mV1 BRT No.: 1 Reagent: 32

WTint : 0 [sec]
Vup : 200 [μl]
Vlow : 10 [μl]
dE : 5 [mV]
dT : 5 [sec]

DL : 1200 [mV/ml]

DetCnt : 10 C1 : (BL-A1)*f*1.269/S

Vmax : 50 [ml] [g/100g]

Vover : 1 [ml]

Reagent name (Reag): Na2S2O3 Equivalent (E): 1 Molarity (M): 0.1 [Mol/I]

Factor (f) : 0.997 Blank (BL) : 47.3915 [ml]

Buret Injection Speed: 500 [ul/sec]